

Department of Biochemistry

Autoclave Testing Log For: January 2022

Autoclave ID	Test Date	Operator	Chemical Indicator Results		Biological Indicator Results		Comments
			Pass	Fail	Pass	Fail	
	1/24	JK					
B175 #1			✓		✓		
#2			✓		✓		
#3			✓		✓		
154 #1			✓		✓		
#2			✓		✓		
220 #1			✓		✓		
#2			✓		✓		
320 #1			✓		✓		
#2			✓		✓		
420			✓		✓		
570			✓		✓		
3406	↓	↓	✓		✓		
2246	1/25	JK	✓		✓		
3246	↓	↓	✓		✓		
ACU	1/25	JK					
B125	↓	↓	✓		✓		
B147	↓	↓	✓		✓		
B2260	↓	↓	✓		✓		