

Department of Biochemistry

Autoclave Testing Log For: June 2021

Autoclave ID	Test Date	Operator	Chemical Indicator Results		Biological Indicator Results		Comments
			Pass	Fail	Pass	Fail	
B175 #1	6/23	SC+EK	✓		✓		
#2			✓		✓		
#3			✓		✓		
154 #1			✓		✓		
#2			✓		✓		
220 #1			✓		✓		
#2			✓		✓		
320 #1			✓		✓		
#2			✓		✓		
420			✓		✓		
570			✓		✓		
3406	6/24	JIK	✓		✓		
2246			✗		✗		Out of Service
3246			✓		✓		
ACU	6/24	JIK					
B125							Out of Service
B147			✓		✓		
B2260			✓		✓		