Department of Biochemistry

Autoclave Testing Log For: November 2021

Autoclave ID	Test Date	Operator	Chemical Indicator Results Pass	Chemical Indicator Results Fail	Biological Indicator Results Pass	Biological Indicator Results Fail	Comments
	11/23	JK					
B175 #1			✓		✓		
#2			✓		✓		
#3			✓		✓		
154 #1			✓		✓		
#2			✓		✓		
220 #1			✓		✓		
#2			✓		✓		
320 #1			✓		✓		
#2			✓		✓		
420			✓		✓		
570			✓		✓		
3406			•		•		Out of Service
2246	11/29	JK	✓		✓		
3246			✓		✓		
ACU							
B125	11/30	JK	✓		✓		
B147	11/29		✓		✓		
B2260	11/30		✓		✓		