Department of Biochemistry

Autoclave Testing Log For: October 2021

Autoclave ID	Test Date	Operator	Chemical Indicator Results Pass	Chemical Indicator Results Fail	Biological Indicator Results Pass	Biological Indicator Results Fail	Comments
	10/27	JK					
B175 #1			✓		✓		
#2			✓		✓		
#3			✓		✓		
154 #1			✓		✓		
#2			✓		✓		
220 #1			✓		✓		
#2			✓		✓		
320 #1							Out of Service
#2			✓		✓		
420			✓		✓		
570			✓		✓		
3406			•				Out of Service
2246			✓		✓		
3246			✓		✓		
ACU	10/28	JK					
B125			✓		✓		
B147			✓		✓		
B2260			✓		✓		